

Figure 25 TAC §289.231(11)(2)

| <u>Machine Type/Type of Use</u>                                       | <u>Years Between Inspections</u>  |
|---|-----------------------------------|
| CT  | 2 years                           |
| Fluoroscopy   | 2 years                           |
| Accelerators, Simulators, and Other<br>Therapeutic Radiation Machines | 2 years                           |
| Radiographic Only   | 3 years                           |
| Podiatric Radiographic Only   | 4 years                           |
| Minimal Threat  | 5 years                           |
| Industrial Radiography  | 1 year                            |
| Other Industrial  | 5 years                           |
| Laser (Human Use/Research/Academic)                                   | As deemed necessary by the agency |
| Other Laser   | As deemed necessary by the agency |
| Mammography   | 1 year                            |

NOTE: The inspection intervals specified above were based upon the average number of health-related violations per inspection by category, as determined from compliance history data. These intervals will be reviewed at least every two years, and appropriate adjustments will be made.